Unpublished Manuscript EU LI Environnement Manuscript T Environment Canada Canada CIN Canada Centre JUL 30 1986 Canadien Centre For Inland **Des Eaux** ORA Waters Intérieures Comparison of Three Microbial Toxicity Screening Tests with the Microtox Test B.J. Dutka and K.K. Kwan

> UNPUBLISHED PEPORT RAPPORT NON PUBLIE

MANUSCRIPT NO. 10-AMD-5-81-BD

Comparison of Three Microbial Toxicity Screening Tests with the Microtox Test

B.J. Dutka and K.K. Kwan

Bull. Environm. Contam. Toxicol.

Comparison of Three Microbial Toxicity Screening Tests with

the Microtox Test

B.J. Dutka and K.K. Kwan

Microbiology Laboratories Section, Analytical Methods Division, National Water Research Institute, Canada Centre

for Inland Waters, Burlington, Ontario L7R 4A6

A variety of test methods, criteria and procedures have been developed to assess the impact of chemical pollutants on aquatic biota. With the increasing awareness of the long term effects of many of those chemical pollutants, research efforts are being directed at short term bioassay tests, in an attempt to quickly alert dischargers as well as monitoring agencies of potential toxic conditions. One of the reasons for the time emphasis is that some effluents may be able to be stopped or contained for short periods for extra treatment, if necessary, but volume problems would make it unrealistic to attempt a 24-hour much less a 96-hour containment. Also by rapid assessment of changes in effluent quality, it may be possible to modify treatment before too great an environmental impact has occurred.

This report details our findings on the following microbiological acute toxicity screening tests; Microtox (developed by Beckman Instruments, Inc.), <u>Spirillum volutans</u>, (Boudre and Krieg, 1974). <u>Pseudomonas fluorescens ATCC-13525</u> (based on the English standard NEN 6509-water-determination of the effect of toxic substances on the growth of a pure culture of bacteria) and <u>Aeromonas hydrophila ATCC 23213</u>, a typical water bacterium (using the P. fluorescens procedure).

MATERIALS AND METHODS

Chemicals: The following chemicals were tested at pH 6.7 (as the sensitivity and stability of the Microtox procedure is based on testing samples where the pH is close to 6.7): Hg^{++} ($HgCl_2$), Zn^{++} (Zn SO₄• 7H₂O), Cu^{++} ($CuSO_4$), Pb^{++} ($PbCl_2$), Ni⁺⁺ (NiCl₂•6H₂O), α -napthol, sodium lauryl sulfate, sodium cyanide, 3,3 dichlorobenzidine, phenol, N-nitrosodiethylamine, dichloromethane, and nitrotriacetic acid.

Toxicity Tests: The Microtox test was performed using the Microtox reagent and following the procedure detailed in the Beckman Instruments Interim Manual No. 110679B-9-80 with a 15 minute incubation time. Spirillum volutans, a large

1 -

aquatic bacterium with a rotating fascicle of flagella at each pole, was used to test the samples for toxicity, following a modification of the procedure developed in 1974 by Bowdre and Krieg (Dutka, 1978). Pseudomonas fluorescens was inocculated into 100 mL of nutrient broth and incubated at 37°C for 16 to 18 h on a rotary shaker (100 rpm). Fifteen mL of this culture was inocculated into 1 L of nutrient broth in a 2 L Erlenmeyer flask which was constantly mixed with a sitrring bar. This was the test inocculum which was dispensed into test flasks within 30 minutes. All chemicals spanned a minimum 4 log concentration gradient encompassing negative and positive effects. To test each chemical, 25 mL of sample was combined with 25 mL of cell inocculum in a, 125 mL Erlenmyer flask, swirled to thoroughly mix and 5 mL of the sample was removed for optical density determinations for time 0. The flask was then placed on a 100 rpm rotary shaker for 18 hours at 37°C. At the end of the incubation period, 5 mL aliquots were once again removed and tested for optical density (650 nm) in a Spectronic 20. All sample concentrations were tested in duplicate. Uninocculated media were used as a negative control and potassium dichromate (0.01 to 50 ppm) was the positive control. The data were graphed and EC50 values established. Similar procedures were followed using a culture of Aeromonas hydrophila, ATCC 23213.

RESULTS AND DISCUSSION

The great variation in sensitivity patterns of the four microbial toxicity screening procedures to the selected chemicals are shown in Table 1. Some of the chemicals have toxicity end point concentrations 100 to 1000 times greater from one test system to the next, e.g., the EC_{50} of Zn^{++} is 3.45 ppm in the Microtox system and 367.0 ppm in the <u>P. florescens</u> test, and sodium lauryl sulfate 1.8 ppm in the Microtox system and 3690 ppm in the <u>A. hydrophila</u> tests. With the exception of Hg⁺⁺, Ni⁺⁺, Pb⁺⁺ and N-Nitrosodiethylamine the 15 minute Microtox procedure (30 to 45 minutes total test) was the most sensitive to the chemicals tested with the <u>P.</u> <u>fluorescens</u> (20 h total test) procedure being the next most sensitive. In no instance were the two h <u>Spirillum volutans</u> test and the <u>A. hydrophila</u> toxicity tests the most sensitive to the chemicals being tested.

Table 1 data clearly illustrates the advantages (speed and sensitivity) of using the Microtox test in toxicity screening procedures. Table 1 is also very supportive of the battery approach for toxicity screening as each procedure is shown to react to different levels of chemicals. Table 1. Sensitivity of Four Acute Toxicity Screening Procedures to Various Chemicals

	Concent	ration in ppm Reaction	Concentration in ppm to Give Typical Endpoint Reaction to Toxicants	ndpoint	
	15-min Microtox EC ₅₀	120-min S. volutans	18 hour P. fluorescens	18 hour A. hydrophila	(J)
Hg++ (Mercury Chloride)	0.046	0.2 b	0.031 a	0.049	1
Zn ⁺⁺ (Zinc Sulfate)	3.45 a	11.6	367.0	500.0 b	
Curr (Copper Sulfate)	3.8 a	10.0	16.75	21.25 b	~
Pb (Lead Chloride)	30.2	40.0	13.8 a	705.0 b	
Ni ^{TT} (Nickel Chloride)	23.0	20•0 b	8.7 a	16.80	
α Naphthol	3.8 a	10.0	>100. b	>100 b	.
Sodium Lauryl Sulfate	1.8 a	43.0	1650.0	3690.0 b	
Sodium Cyanide pH 6.7	2.76 a	83.0	14.0	25.25 b	~~~
3,3 Dichlorobenzidine	0.058 a	- 16.0	>100 b	>100 b	
Phenol	34.3 a	300.0	875.0	1620.0 b	
N-Nitrosodiethylamine	138.0 a	570.0	>5000 b	>5000 b	
Dichloromethane	>1000	>1600	>10000	>10000 b	_
Nitrotriacetic Acid	>1000	>800 -	719.0 a	1500.0 b	
					7

a = most sensitive; b = most resistant

Ŀ

3 -

The Microtox toxicity testing procedure, which is now undergoing a very thorough review by many laboratories, has shown to have some problems with reproducibility. For instance, in Table 2, six toxicants are compared with only two substrates (Hg¹⁺¹ and phenol) showing similar results by two laboratories (Beckman Instruments, Inc., and ours). Also, in a study (reported by Beckman Instruments, Inc.) using sodium pentachlorophenate, 30 separate assays were performed using 30 separate vials of Microtox reagent and recording 5 minute EC_{50} and 15 minute EC_{50} results, it was found: (a) 5 minute EC_{50} concentrations varied from 0.574 to 0.375 ppm (65.3%); (b) the 15 minute EC_{50} concentrations varied from 0.425 to 0.275 ppm (64.7%); and (c) the ratio's between 5 and 15 minute EC_{50} readings varied from 1.27 to 1.45.

	5-Minute EC ₅₀ (ppm)		
Toxicant	Bulich <u>et</u> al.*	Duluth EPA**	Dutka-Kwan
Hg++	0.065	0.052	0.064
Sodium lauryl sulfate	1.6	-	3.19
Zn ⁺⁺ Cu ⁺⁺	2.5	52.0	13.8
Cu ⁺⁺	8.0	15.1	19.5
Phenol	25.0	40.7	28.0
Ethanol	31,000	56,706	-

TABLE 2. COMPARISON OF MICROTOX EC₅₀ VALUES OBTAINED IN THREE LABORATORIES

* A.A. Bulich, M.W. Greene and D.L. Isenberg (1980).

** Data produced by Miss Carolann Curtis, U.S. EPA, Environmental Research Laboratory, Duluth, Minnesota 55804

Thus, there is a problem of reproducibility of data within one laboratory, and also between laboratories (Table 2), probably related to variations in the cell suspension. Another example of this type of reproducibility problem is shown with EC_{50} values for ethanol. The 5 minute EC_{50} value obtained for ethanol by the Beckman Instruments Laboratory was 31,000 ppm (Bulich et al., 1980) and by Chang et al., (1981), 47,000 ppm. However, LC_{50} values in fish toxicity tests are also known to show similar reproducibility problems.

Reviewing the data from the four toxicity assessing techniques, it is obvious that each procedure has its own toxicity sensitivity pattern and cannot be readily correlated with the other procedures. There are areas of concurrence as well as areas of wide divergence in sensitivity to toxicants. Table 3 contains data from Ryssov-Nielsen's (1975) study which used TTC-dehydrogenase and short term Warburg tests for assessing toxicity and illustrates the variety of substrate concentrations that ellicit typical end points.

Toxic Substrate	^l TCC 50%* Inhibition Test (ppm)	^l Warburg 50%* Inhibition Test (ppm)	Microtox 15-Min EC ₅₀ (ppm)
Zn	26	1.4	3.45
Hg	1.5	0.6	0.046
Cn	0.47	4.7	2.76
	<u>S. volutans</u> 90% Inhibition 2-hour (ppm)	P. fluorescens EC50 (ppm)	A. hydrophila EC ₅₀ (ppm)
Zn	11.6	367.0	500.0
Hg	0.2	0.031	0.049
Cn	83.0	14.0	25.25

TABLE 3. COMPARISON OF TOXICITY OF SELECTED CHEMICALS BY SIX MICROBIAL TOXICITY TESTING TECHNIQUES

¹ H. Ryssov-Nielsen (1975).

Similarly, in a U.S. EPA sponsored project (EPA Quality Assurance Newsletter, Vol. 4:2, April, 1981) of an effluent study comparing 24 hour fathead minnow and Daphnia pulex LC₅₀ tests with the 5 minute Microtox test, it was found that Microtox indicated the presence of toxicity in 81% of the effluents that were toxic to the fathead minnows. The Microtox test indicated the presence of only 62% of the samples which were toxic to Daphnia pulex. From the above, it would appear there is some overlap or "correct" guesses where all systems indicate a positive effect, but no system was able to predict the 100% presence of toxicant to another species.

From Tables 1 and 2 and the above EPA sponsored data, it is very obvious that no single biological testing procedure can predict the presence of all toxicants which might effect aquatic organisms or be eventually bioaccumulated and affect their predators or man.

In spite of the above, there is no doubt that the Microtox system is a sensitive toxicity assaying procedure which has as its major benefit, a quick turnaround time which makes it an ideal member of a battery of screening tests. Used alone, we believe the Microtox system may have its most useful application in the monitoring of a supposedly consistent effluent stream. Thus, any deviations from the established norm could be easily and quickly noted and rectified. One major drawback of the Microtox test maybe its inability to test some samples at their natural pH.

REFERENCES

DUTKA, B.J. Methods for Microbiological Analysis of Waters, Wastewaters, and Sediments. Dept. of Fish. and Environ., IWD, CCIW, BURLINGTON, ONTARIO. (1978).

RYSSOV-NIELSEN, H. Water Res., 9:1179. (1975).

- U.S. EPA QUALITY ASSURANCE NEWSLETTER, Vol. 4, No. 2, April, (1981).
- BECKMAN INSTRUMENTS, INC. Interim Manual No. 110679B-9-80. (1980).

BOWDRE, J.H. and N.R. KRIEG. Bulletin 69, VWRRC, Virginia Polytechnic Institute and State University, Blacksburg, Virgina, 24061. (1974).

BULICH, A.A., M.W. GREENE, and D.L. ISENBERG. Fourth Ann. Symp. on Aqua. Toxicol. ASTM. (1980).

CHANG, J.C., P.B. TAYLOR and F.R. LEACH. Environ. Contam. Toxicol. 26, 150. (1981).

