

# Class II application content and classification guidance



**Health Canada is the federal department responsible for helping the people of Canada maintain and improve their health.** Health Canada is committed to improving the lives of all of Canada's people and to making this country's population among the healthiest in the world as measured by longevity, lifestyle and effective use of the public health care system.

Également disponible en français sous le titre :  
Lignes directrices sur le contenu et la classification des demandes d'homologation d'instruments de classe II

To obtain additional information, please contact:

Health Canada  
Address Locator 0900C2  
Ottawa, ON K1A 0K9  
Tel.: 613-957-2991  
Toll free: 1-866-225-0709  
Fax: 613-941-5366  
TTY: 1-800-465-7735  
E-mail: [publications-publications@hc-sc.gc.ca](mailto:publications-publications@hc-sc.gc.ca)

© His Majesty the King in Right of Canada, as represented by the Minister of Health, 2025

Publication date: November 2025

Information contained in this publication or product may be reproduced, in whole or in part, and by any means, for personal or public non-commercial purposes without charge or further permission, unless otherwise specified. Commercial reproduction and distribution are prohibited except with written permission from Health Canada. To obtain permission to reproduce any content owned by the Government of Canada available for commercial purposes, please contact [pubsadmin@hc-sc.gc.ca](mailto:pubsadmin@hc-sc.gc.ca).

Cat.: H164-403/2025E-PDF  
ISBN: 978-0-660-79518-8  
Pub.: 250298

## Table of Contents

<b>Chapter 1: Regional administrative</b> .....	<b>4</b>
<b>1.01 Cover Letter</b> .....	<b>4</b>
<b>1.04 Application form and administrative information</b> .....	<b>5</b>
<b>1.05 Listing of devices</b> .....	<b>6</b>
<b>1.06 Quality management system, full quality system or other regulatory certificates</b> .....	<b>7</b>
<b>1.09 User fees</b> .....	<b>7</b>
<b>1.10 Pre-submission correspondence and previous regulator interactions</b> .....	<b>8</b>
<b>Chapter 2: Submission context</b> .....	<b>9</b>
<b>2.04 Device description</b> .....	<b>9</b>
<b>Chapter 5: Labelling and promotional material</b> .....	<b>10</b>
<b>5.02 Product/package labels</b> .....	<b>11</b>
<b>5.03 Package insert and instructions for use</b> .....	<b>12</b>
<b>5.04 E-labelling</b> .....	<b>12</b>
<b>5.05 Healthcare professional labelling</b> .....	<b>14</b>
<b>5.06 Patient labelling</b> .....	<b>14</b>
<b>5.07 Technical and/or operators manual</b> .....	<b>14</b>
<b>5.09 Product brochures</b> .....	<b>15</b>
<b>5.10 Other labelling and promotional material</b> .....	<b>15</b>

## Chapter 1: Regional administrative

Folder name: 1-REGADMIN

Note: No files or content should be included at this level. Only sub-folders of this folder should contain documents.

### 1.01 Cover Letter

Folder name: 1.01-CoverLetter

#### IMDRF common content

- a) The cover letter should state applicant or sponsor name and/or their authorized representative, the type of submission, the common name of the device (if applicable), device trade name or proprietary name (both of the base device and a new name if one is given to the new version/model of the device) and include the purpose of the application, including any changes being made to existing approvals
- b) If applicable and accepted by the regulator, it should include information pertaining to any master files referenced by the submission
- c) If applicable, acknowledgement that a device sample has been submitted or offered alternatives to allow the regulator to view or access the device (when the regulator requests a sample)
- d) If the submission is requesting approval of a change that is the result of CAPA due to a recall, this should be stated
- e) If the submission is in response to a request for information from the regulator this should be stated and the date of that letter should be included as well as any reference number(s)
- f) If the submission is unsolicited information (where accepted), this should be stated and any related reference number(s) provided
- g) Identification of the regulatory jurisdiction(s) in which marketing is sought

**Note:** The cover letter should not contain any detailed scientific information.

#### Health Canada guidance

Any information submitted to Health Canada should be accompanied by a cover letter. The cover letter should include the purpose of the application and a brief description of the package being submitted. It may also include information pertaining to Proprietary Information Submission. When applicable, identify the licence/application number and section associated with the submission (for example, Section 36 or Section 39 of the Medical Devices Regulations).

When responding to a Screening Deficiency Notice, applicants should provide their response in a question-and-answer format accompanied by a copy of this notice. This information is to be filed within folder 1.01 – Cover Letter.

## Classification

New and amendment applications: Required

### 1.04 Application form and administrative information

Folder name: 1.04-ApplicationFormAdministrativeInfo

#### IMDRF Health Canada content

Health Canada application forms should be included here.

#### Health Canada guidance

Obtain the appropriate [medical device licence application form](#). For more information on how to complete the application form for a medical device licence, consult:

- [Guidance on how to complete the application for a new medical device licence](#)

For applications submitted through the Regulatory Enrolment Process (REP), the medical device licence application forms are not required. Instead, applicants must complete the Regulatory Transaction (RT) template and, if applicable, the Application Information (AI) template, which generate the required REP Extensible Markup Language (XML) files. These XML files replace the standard application forms and must be included in the regulatory transaction. Do not submit the standard application forms, as the information is already captured within the XML files. For more information, consult:

- [Regulatory enrolment process \(REP\)](#)

## Classification

New and amendment applications: Required

### 1.05 Listing of devices

Folder name: 1.05-ListingofDevice(s)

#### IMDRF common content

A table listing each variant/model/configuration/component/accessory that is the subject of the submission and the following information for each variant/model:

- a) the identifier (for example, bar code, catalogue, model or part number, UDI)
- b) a statement of its name/description that provides (for example, Trade name, size, material)

#### Notes:

- i. A model/variant/configuration/component/accessory of a device has common specifications, performance and composition, within limits set by the applicant.
- ii. Typically each item listed should be available for sale. For example, if everything is sold as part of a kit, then this list would only include the kit. You do not need to list all components that may be sold within a kit/set, unless the component is available for sale independently of the kit.

#### Health Canada guidance

The listing of devices is part of the medical device licence application form. A separate file in this folder is not required for submissions sent on physical media or by email.

For applications that are submitted through the REP, applicants should complete the Device Details spreadsheet found in the [REP for medical devices page](#).

Please ensure that there are no additional security settings applied to submitted files. This includes password protection, restricted access, Digital Rights Management (DRM), or Information Rights Management (IRM). Files with these settings may be blocked by Health Canada's security policies, preventing these documents from being accessed.

## Classification

New applications: Conditionally required – If application is submitted through REP

Amendment applications: Conditionally required – If application impacts the manufacturer's device identifier listing and is submitted through REP

## **1.06 Quality management system, full quality system or other regulatory certificates**

Folder name: 1.06-QMSFullQSorOtherRegulatoryCerts

### **IMDRF Health Canada content**

Health Canada will only accept MDSAP certificates that have been issued by recognized auditing organizations.

### **Health Canada guidance**

A copy of the quality management system certificate should be included here if applicable.

### **Classification**

New licence applications: Required

Amendment licence applications: Conditionally required – Include the quality management certificate if the amendment includes a change in the manufacturer's name.

## **1.09 User fees**

Folder name: 1.09-UserFees

### **IMDRF Health Canada content**

Health Canada user fee forms should be included here.

### **Health Canada guidance**

When submitting an application for a medical device licence, include the [medical device licence application fee form](#) to avoid delays in the processing of your application. Guidance on fees for the review of medical device licence applications can be found in the [Guidance Document: Fees for the Review of Medical Device Licence Applications](#).

## Classification

New and amendment applications: Required

### 1.10 Pre-submission correspondence and previous regulator interactions

Folder name: 1.10-Pre-SubmissionCorrespondence-PreviousRegulatorInteractions

#### IMDRF common content

- a) During the product lifecycle, pre-submission correspondence, including teleconferences or meetings, may be held between the regulator and the applicant. Further, the specific subject device may have been subject to previous regulatory submissions to the regulator. The contents should be limited to the subject device as similar devices are addressed in other areas of the submission. If applicable, the following elements should be provided:
- i. List prior submission or pre-submissions where regulator feedback was provided
  - ii. Prior submissions should include identification of applicable submission number
  - iii. For any pre-submission activities that have not previously been assigned any tracking/reference number, include the information package that is submitted prior to pre-submission meetings, the meeting agenda, any presentation slides, final meeting minutes, responses to any action items arising from the meetings, and any email correspondence related to specific aspects of the application
  - iv. Issues identified by the regulator in prior submissions (that is, clinical study applications, withdrawn/deleted/denied regulatory submission) for the subject device
  - v. Issues identified and advice provided by the regulator in pre-submission interactions between the regulator and the applicant/sponsor
  - vi. Explain how and where the prior advice was addressed within the submission

Or

b) Affirmatively state there has been no prior submissions and/or pre-submission interactions for the specific device that is the subject of the current submission

**Note:** The scope of this section is limited to the particular regulator to which the submission is being submitted (for example Health Canada does not need pre-submission information relating to interactions with ANVISA).

### **Classification**

New and amendment applications: Conditionally required – If applicable

## **Chapter 2: Submission context**

Folder name: 2-CONTEXT

Note: No files or content should be included at this level. Only sub-folders of this folder should contain documents.

### **2.04 Device description**

Folder name: 2.04-DeviceDescription

#### **IMDRF common content**

No content at this level.

#### **Classification**

New licence applications: Optional but recommended

Amendment licence applications: Optional but recommended

#### **2.04.04 Reference and comparison to similar and/or previous generations of the device (nIVD)**

Folder name: 2.04.04-Ref-ComparisontoSimilarand-orPreviousGen

#### **2.04.05 Reference and comparison to similar and/or previous generations of the device (IVD)**

Folder name: 2.04.05-Ref-ComparisontoSimilarand-orPreviousGen

### **IMDRF common content**

- a) A list of similar devices (available on local and international market) and/or previous generation of the devices (if existent) relevant to the submission. This should include any similar/previous generation devices that were previously reviewed and refused by the subject regulator
- b) Description of why they were selected
- c) A key specification comparison, preferably in a table, between the references (similar and/or previous generation) considered and the device

### **IMDRF Health Canada content**

- a) If the application is an amendment to a licensed device or is based on a modification of a licensed device, a description of the modifications is required (for example, changes in design, performance, and indications).
- b) Comparisons can be used to support the safety and effectiveness of the device if they are made to a currently licensed device in Canada. If this method is used, ensure the Canadian Medical Device Licence Number of the comparator is stated.

### **Health Canada guidance**

When adding a device identifier, comparative information is necessary to validate compliance with licence type requirements. Include a comparison table that outlines all differences between the devices represented by the new identifiers and a closely related licensed comparator(s). For each difference identified, include information that explains why the new identifiers are part of the licence's designated licence type.

### **Classification**

New licence applications: Optional

Amendment licence applications: Conditionally required - when the change is to add a device identifier

## **Chapter 5: Labelling and promotional material**

Folder name: 5-LABELLING

Note: No files or content should be included at this level. Only sub-folders of this folder should contain documents.

## 5.02 Product/package labels

Folder name: 5.02-Product-PackageLabels

### IMDRF common content

Samples of the primary and secondary packaging labels.

#### Notes:

- i. Do not include shipping labels.
- ii. The sponsor/applicant should explicitly address any existing regional regulatory guidance related to labelling the subject nIVD medical device.

### IMDRF Health Canada content

- a) All labelling must comply with sections 21 to 23 of the Medical Devices Regulations.
- b) Consult the [guidance for the labelling of medical devices, not including in vitro diagnostic devices](#) or the [guidance for the labelling of in vitro diagnostic devices](#) as applicable.

### Health Canada guidance

Include samples of the primary and secondary packaging labels but exclusive of labels for shipping.

#### Notes:

- i. All labelling must be provided in English or French, both official languages are to be available upon request.
- ii. Labelling for near-patient devices must also be provided in French and English.

### Classification

New and amendment applications: Required

### 5.03 Package insert and instructions for use

Folder name: 5.03-PackageInsert-InstructionsforUse

#### IMDRF common content

Package Insert/Instructions for Use included in the package, when required or provide support for why this element is not applicable.

Note: The sponsor/applicant should explicitly address any existing regional regulatory guidance related to labelling the subject device.

#### IMDRF Health Canada content

- a) All labelling must comply with sections 21 to 23 of the Medical Devices Regulations.
- b) Consult the guidance for the [labelling of medical devices, not including in vitro diagnostic devices](#) or the [guidance for the labelling of in vitro diagnostic devices](#) as applicable.
- c) The current version and date of the instructions for use must be stated.

#### Health Canada guidance

The package insert or instructions for use must be included when required. If this element is not applicable, justification should be provided for its exclusion.

- a) All labelling must be provided in English or French, both official languages are to be available upon request.
- b) Labelling for near-patient devices must also be provided in French and English.
- c) For IVD devices, package inserts must include a summary of the performance specifications, in addition to all other relevant information.

#### Classification

New and amendment applications: Conditionally required – If applicable

### 5.04 E-labelling

Folder Name: 5.04-e-labelling

### **IMDRF common content**

In addition to the e-labelling itself, the following should be provided:

- a) For eligible medical devices and Software as a Medical Device, the applicant needs to identify which form of e-labelling is being used (for example electronic storage system or built-in system, website)
- b) Details of risk management in relation to e-labelling. If this is part of the overall risk management, refer to it here
- c) When IFUs are requested, a description of the procedure and operations on providing IFU's when requested
- d) Written information for users on the webpage identifying where the IFU and further information can be found in relevant languages
- e) A description on how the e-labelling requirements for the website have been met
- f) If a video/app is available to demonstrate how the device or test is intended to function or perform and be interpreted (for IVD devices). Provide a link as well as details about how it is maintained and updated throughout the life cycle of the device

### **IMDRF Health Canada content**

For devices that are not sold to the general public, IFUs may be provided as downloadable from the internet and/or on electronic data storage devices, for example compact disc, digital video disc, USB flash drive, etc. The electronic label or URL must accompany the device at the time of sale and/or delivery and be displayed in a manner that alerts the user to its purpose. A Letter of Attestation must also be included with the application. Refer to the [guidance for the labelling of medical devices, not including in vitro diagnostic devices](#) or the [guidance for the labelling of in vitro diagnostic devices](#), for additional information.

If a video/app is available as described in f) above, the video should be available in both French and English.

### **Classification**

New and amendment applications: Optional

## **5.05 Healthcare professional labelling**

Folder name: 5.05-HCPLabelling

### **IMDRF common content**

Labelling directed at the healthcare professional other than the package insert, such as the surgical manual.

### **Classification**

New and amendment applications: Conditionally required - If applicable

## **5.06 Patient labelling**

Folder name: 5.06-PatientLabelling

### **IMDRF common content**

Labelling directed at the patient other than the package insert, such as informational material written to be comprehended by the patient or lay caregiver.

### **Classification**

New and amendment applications: Conditionally required - If applicable

## **5.07 Technical and/or operators manual**

Folder name: 5.07-Technical-OperatorManual

### **IMDRF common content**

Labelling directed at the technical users and operators of medical devices focusing on the proper use and maintenance of the device and surgical technique instructions.

### **Classification**

New and amendment applications: Conditionally required - If applicable

## 5.09 Product brochures

Folder name: 5.09-ProductBrochures

### IMDRF Health Canada content

Draft product brochures available at the time of application

#### Classification

New and amendment applications: Required

## 5.10 Other labelling and promotional material

Folder Name: 5.10-OtherLabelling-PromotionalMaterial

### IMDRF common content

Heading for other information that may be important to the submission but that does not fit in any of the other headings of this chapter.

Individual jurisdictions may have their own regulations or requirements regarding other labelling elements or advertising and promotional materials. If necessary, this section can be used to address jurisdiction-specific regulations or requirements involving other labelling elements other than those described elsewhere in this section, including advertising and promotional materials.

### Health Canada guidance

Include any other relevant labelling or promotional material that does not fit in any other folder.

#### Classification

New and amendment applications: Conditionally required - If applicable